



# MILESTONE INSTRUMENTS

WZ-8 D DAYAL SAR COLONY, UTTAM NAGAR WEST, N.DELHI-110059 INDIA  
Email: - info@milestoneinstruments.com, milestoneinstrumentsindia@gmail.com  
Website:-www.milestoneinstruments.com, www.milestoneequipments.com  
www.geotextiletesting.com ☎ +91 98911 99011, +91 11 2533 2422

## **MI.GE-1012 Creep tester**

It consists of a loading frame for the specimen to be mounted, the load to be correctly applied, and strain monitored. The connections between jaws and loading apparatus have sufficient freedom that the load is applied uniformly across the width of the specimen. The jaws shall grip the specimen with sufficient firmness to allow minimum slippage

### **Technical Specifications:**

Maximum Load	5 kg per sample
Sample holder	3 number
Sample size	20×5 cm
Max. Displacement	150 mm
Displacement accuracy	1 mm
Related Standards	IS-14739