



MILESTONE INSTRUMENTS

WZ-8 D DAYAL SAR COLONY UTTAM NAGER DELHI-110059 INDIA
Email:- info@milestoneinstruments.com, milestoneinstrumentsindia@gmail.com
Website:-www.milestoneinstruments.com
☎ +91 9891199011, +91 11 25332422

MI.GE-1008 Index Puncture Test (Ball Bursting Strength of Textiles)

The geotextile specimen is clamped to obtain an inside diameter of 44.45 mm and force is applied through steel ball having 25 mm diameter at the rate of 305 mm/min and continue at that speed until the specimen bursts. Record to the nearest 0.5 N the ball-bursting strength of the specimen..

Technical Specifications:

Inside diameter	44.45 mm Dia meter
Loading Plunger Size	25 MM Steel Ball
Speed of the Test	305 mm/min
Related Standards	ASTM D 3787, ASTM D6241, ASTM 3786.